



Certificate of Analysis

Product Compact Dry PA 1400plates (Code 54062)

Lot No 028106

Expiry Date 2022 . 12

	Specification	Results
Appearance	Light pink. No particular matters.	Pass
Loss of drying	Less than 10.0%	4 . 9 %
Aseptic Test	Gamma irradiation over 16kGy	Pass
pH	6.8 ~ 7.6	7 . 4

Evaluation of Performance Specification

- ① When following test strains are inoculated to incubate at 35°C for 45 - 51 hours, good growths and color developments should be observed.

<u>Test strain</u>	<u>Growth</u>	<u>Color development</u>
<i>Pseudomonas aeruginosa</i> ATCC 9027	+	red
<i>Pseudomonas aeruginosa</i> ATCC 27853	+	red
<i>Pseudomonas aeruginosa</i> ATCC 10145	+	red

- ② When following test strains are inoculated to incubate at 35°C for 20 - 24 hours, growths should be inhibited.

<u>Test strain</u>
<i>Escherichia coli</i> ATCC 8739
<i>Bacillus subtilis</i> ATCC 6633
<i>Staphylococcus aureus</i> ATCC 6538

Judge: Acceptable quality.

Certificate issue date : 2021 . 6 . 24

Signed (QC Manager)

Emiko Takahashi